


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>Accredited to ISO/IEC 17025:2017</p>	<b>Jet-Care International Inc</b>  <b>Issue No: 031    Issue date: 29 July 2022</b>	
	<b>3 Saddle Road</b> <b>Cedar Knolls</b> <b>New Jersey</b> <b>07927</b> <b>USA</b>	<b>Contact: Mrs A Potere</b> <b>Tel: + 001-973-292-9597</b> <b>E-Mail: alisonpotere@jet-care.com</b> <b>Website: www.jet-care.com</b>
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<b>COOLANTS</b>  Engine Coolants (monitoring to determine fitness for purpose)	<u>Chemical and Physical Tests</u>  Anions: Chloride, Nitrate, Nitrite, Phosphate Sulphate  Determination of Coolant Properties – by Refractive Index with calculation of freeze point, %ethylene glycol and boiling point	Documented In-House Methods  M054 based on ASTM D5827 by Ion Chromatography  M052 based on ASTM D3321
<b>PETROLEUM and PETROLEUM PRODUCTS, OILS and LUBRICANTS</b>  Oils	<u>Chemical and Physical Tests</u>  Blotter Spot Test  Colour  Crackle test  Density and Specific Gravity  Elemental analysis: Ag, Al, B, Ba, Be, Ca, Cd, Cl, Cr, Cu, Fe, Li, K, Mg, Mn, Mo, Na, Ni, P, Pb, S, Sb, Si, Sn, Ti, V, W, Zn  Elemental Analysis: Ag, Al, B, Ba, Ca, Cr, Cu, Fe, Mg, Mn, Mo, Ni, P, K, Na, Pb, S, Si, Sn, Ti, V, Zn.	Documented In-House Methods  ASTM D7899 (modified)  M009 based on ASTM D6045 Automatic Tristimulus method with conversion to ASTM D1500  M026  M011 based on ASTM D5002  M019 based on ASTM D5185 by ICP  ASTM D5185 by ICP AES



1961  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Jet-Care International Inc**  
**Issue No: 031 Issue date: 29 July 2022**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM and PETROLEUM PRODUCTS, OILS and LUBRICANTS (cont'd)	<u>Chemical and Physical Tests</u> (cont'd)	Documented In-House Methods (cont'd)
Oils (cont'd)	Elemental Analysis: Ba, B, Ca, Cu, Mg, Mo, P, S, Zn.	ASTM D4951 by ICP AES
	Ferromagnetic particles	ASTM D8184 by Particle Quantifier
	Flash Point (Cleveland Cup)	M020 based on ASTM D92
	Flash point Go-no-go	M001 Closed cup based on ASTM D7094 and ASTM D93A (mode) Eralytics Automatic Flash Point tester
	Fuel Dilution	M037 by GC based on ASTM D3524
	General filter debris	M032
	Initial pH (IpH)	M016 by IpH meter
	Insoluble material	M010 by membrane filtration
	Measurement of Antioxidant Content	M039 based on ASTM D6971 by RULER
	Metallic debris	M022
	Oxidation, Nitration, Soot, etc	M017 using FTIR
	Particle (alloy type)	M013 by Scanning Electron Microscope
	Size and Number of Particles	M033 based on NAS 1638, ISO4406, SAE4059
	Total Acid Number	M007 based on ASTM D664
	Total Base Number	M006 based on ASTM D2896
	Total Base Number	ASTM D2896 by titration
	Total Base Number (0.1 to 250mg KOH/g)	ASTM D4739-17 by titration



1961  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Jet-Care International Inc**  
**Issue No: 031 Issue date: 29 July 2022**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM and PETROLEUM PRODUCTS, OILS and LUBRICANTS (cont'd)	<u>Chemical and Physical Tests</u> (cont'd)	Documented In-House Methods (cont'd)
Oils (cont'd)	Viscosity	M002 based on ASTM D445
	Viscosity	ASTM D445
	Viscosity Index calculated from Viscosity at 40°C and 100°C	ASTM D2270-10 (2016) by calculation
	Viscosity between -10°C and -35°C	ASTM D5293 using Cold Cranking simulator
	Low Temperature Viscosity	ASTM D2983 by Rotational Viscometer (Brookfield Viscosity)
	Nitrogen	ASTM D5291 by Instrumental Analyser
	Sulfur	ASTM D2622 by Wavelength Dispersive X-ray Fluorescence Spectrometry
	Water content	M023 based on ASTM D6304
TRANSFORMER OILS		Documented In-House Methods
	Dissolved Gases	M050 based on ASTM D3612 by Head Space Gas Chromatography
	Dielectric Breakdown Voltage	M051 based on ASTM D1816 and ASTM D877

END