

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0261

Accredited to  
ISO/IEC 17025:2017

### SPECTRO

a trading name of Palace International Limited

Issue No: 036 Issue date: 09 March 2021

Hatchwood Place  
Farnham Road  
Odiham  
Hampshire  
RG29 1AB

Contact: Mr R Medhurst  
Tel: +44 (0)1256 704000  
Fax: +44 (0)1256 704006  
E-Mail: enquiries@spectro-oil.com  
Website: www.spectro-oil.com

Testing performed at the above address only

#### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM and PETROLEUM PRODUCTS OIL, LUBRICANTS, DEBRIS AND HYDRAULIC FLUID	<u>Chemical and Physical Tests</u>	Documented In-House Methods in the series M 000 as listed below
	Air release value of lubricating and hydraulic oils - Range 0 - 30 mins	M 028 based on IP 313 and ASTM D3427
	Antioxidant content	In house method M039 based on ASTM D6971 by RULER
	Apparent Viscosity (Cold Cranking Simulator) at -10, -15, -20, -25, -30 and -35 °C of Oils and Lubricants	ASTM D5293-17a
	Blotter Spot test	ASTM D7899-13 (modified)
	Colour (Lovibond) - Range 0 - 8 Lovibond units	M 009 based on IP 196, ISO 2049 and ASTM 1500
	Conductivity	M 021 based on ASTM D2624 and IP 274
	Debris content extracted from filter elements - Range 0 - 100% wt	M 032 by filtration/SEM
	Density and Specific Gravity - Range 0.5 - 1.5 g/ml	M 011 based on ASTM D5002
Elemental analysis of dark oils - The following elements within spectral range 130 to 800 nm Ag, Al, B, Ba, Ca, Cd, Cl, Cr, Cu, Fe, K, Li, Mg, Mn, Mo, Na, Ni, P, Pb, S, Sb, Si, Sn, Ti, V, Zn	M 019 based on ASTM D5185 By ICP-AES	



0261  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SPECTRO**  
**a trading name of Palace International Limited**  
**Issue No: 036 Issue date: 09 March 2021**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM and PETROLEUM PRODUCTS OIL, LUBRICANTS, DEBRIS AND HYDRAULIC FLUID (cont'd)	<u>Chemical and Physical Tests</u> (cont'd)	Documented In-House Methods in the series M 000 as listed below
	Elemental analysis of Aviation oils - The following elements within spectral range 130 to 800 nm; Ag, Al, B, Be, Ca, Cl, Cr, Cu, Fe, K, Li, Mg, Mn, Mo, Na, Ni, P, Pb, S, Sb, Si, Sn, Ti, V, W, Zn	M 019 AES based on ASTM D5185 By ICP-AES
	Elemental analysis of Oils Ag, Al, Ba, B, Ca, Cr, Cu, Fe, K, Mg, Mn, Mo, Na, Ni, P, Pb, S, Si, Sn, Ti, V, Zn	ASTM D5185-18 By ICP-AES
	Elemental Analysis of additive elements in Lubricating Oil Ba, B, Ca, Cu, Mg, Mo, P, S, Zn	ASTM D4951-14 By ICP-AES
	Evaporation Test (NOACK) of oils and lubricants	CEC-L-40-93
	Ferromagnetic particles in oil - Range 15 - 750 PQ units	M 025 by Particle Quantifier
	Flash Point - Range Ambient - 300 °C	M 001 by Go, No-go based on IP 303:Part 1 (obsolete) by Setaflash or based on IP303 / ASTM D7094 Eralytics Automatic Closed cup
	- Range 80 - 300 °C	M 020 by Cleveland open cup based on ASTM D92 and IP 36 by Cleveland Open Cup
	- Range 40 - 360 °C	M 031 by Pensky Marten closed cup based on IP 34 and ASTM D93
	Foaming characteristics of lubricants - Range 0 ml (nil) to <1000 ml	M 027 based on IP 146 and ASTM D892 using Seta Dual-twin foam bath



0261  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SPECTRO**  
**a trading name of Palace International Limited**  
**Issue No: 036 Issue date: 09 March 2021**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM and PETROLEUM PRODUCTS OIL, LUBRICANTS, DEBRIS AND HYDRAULIC FLUID (cont'd)	<u>Chemical and Physical Tests</u> (cont'd)	Documented In-House Methods in the series M 000 as listed below
	HTHS (Dynamic Viscometry) at 100 and 150°C of Oils and Lubricants	CEC L-36-90 by Ravenfield High Shear Rate Tapered Plug viscometer
	Insoluble Matter: - heptane or toluene Range 0.01 - 20 % w/w	M 003 by High Speed Centrifuge based on IP 316
	Pentane - Range 0.01 - 5 % w/w	M 010 by Membrane Filtration based on MM 1068
	Metallic debris	M 022 by Chemical analysis
	Micro carbon residue - Range 0.10 % - 30.0 % m/m	M 030 based on ASTM D4530
	Neutralization number: Acid - Range 0.01 - 20.0	M 034 by colour titration based on IP 139 and ASTM D974
	Nitrogen in Oils and Lubricants	ASTM D5762-18a by combustion and chemiluminescence
	Nitrogen in Oils and Lubricants	ASTM D5291-16 by Nitrogen analyser
	TAN Range 0.1 - 20 mg KOH/g	M 007 based on ASTM D664 and IP 177



0261  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SPECTRO**  
**a trading name of Palace International Limited**  
Issue No: 036 Issue date: 09 March 2021

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM and PETROLEUM PRODUCTS OIL, LUBRICANTS, DEBRIS AND HYDRAULIC FLUID (cont'd)	<p><u>Chemical and Physical Tests</u> (cont'd)</p> <p>Total Base Number (TBN) - Range 0.1 - 60 mg KOH/g</p> <p>- Range 0.1-60 mg KOH/g</p> <p>- Range 0.1-300 mg KOH/g</p> <p>Oxidation, nitration, sulphates, glycol, water</p> <p>spectral range 7400 - 375 cm<sup>-1</sup> - Glycol POS/NEG - Oxidation 0 - 50 (abs/cm) - Nitration 0 - 50 (abs/cm) - Water 0 - 5 (% by wt) - Soot 0 - 2 (% by wt)</p> <p>Particle examination - qualitative identification of alloy Type</p> <p>Size and number of particles Range 5 - 100 microns</p> <p>Initial pH Range 0-14 pH units</p> <p>Sulphated Ash in Oils and Lubricants</p> <p>Sulphur</p>	<p>Documented In-House Methods in the series M 000 as listed below</p> <p>M 005 based on ASTM D664 and IP 177</p> <p>M 006 based on IP 276 and ASTM D2896-11</p> <p>ASTM D2896-15 by Metrohm titrator</p> <p>M 017 by FTIR</p> <p>M 013 by Scanning Electron Microscope</p> <p>M 033 based on NAS1638, ISO4406, SAE AS4059</p> <p>M 016 based on ASTM D7946</p> <p>ASTM D874-13a</p> <p>ASTM D2622-16 by WDXRF</p>



0261  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**SPECTRO**  
**a trading name of Palace International Limited**  
Issue No: 036 Issue date: 09 March 2021

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM and PETROLEUM PRODUCTS OIL, LUBRICANTS, DEBRIS AND HYDRAULIC FLUID (cont'd)	<u>Chemical and Physical Tests</u> (cont'd)	Documented In-House Methods in the series M 000 as listed below
	Viscosity Range 1-1000 cSt	M 002 based on IP 71, Section 1, by Semi-Automatic Viscometer Method, ASTM D445-19
	Viscosity at 40°C and 100°C	ASTM D445-19 By Manual and automated viscometers
	Viscosity Index calculated from viscosity results at 40°C and 100°C	ASTM D2270-10 (2016)
	Water Content - Range 0.001 - 10% or 10 - 100,000 ppm	M 023 by automatic Karl Fischer based on ASTM D6304c
Insulating Liquids	Water contamination, positive or negative - crackle test	M 026 by Seta Flash Go, No-go flashpoint tester
Insulating Liquids	Dielectric Breakdown Voltage	ASTM D1816 using VDE electrodes
Insulating Liquids	Water content	ASTM D1533 by Coulometric Karl Fischer
Electrical Insulating Oils	Dissolved Gases	ASTM D3612 by Gas Chromatography
Automatic Transmission Fluids, Hydraulic Fluids, and Lubricants	Low temperature Viscosity	ASTM D2983 by Rotational Viscometer
END		